

Controller DULCOMETER® D1Cb/D1Cc

The water analysis workhorse

ProMinent®



The D1Cb/D1Cc controller is a 1-channel P/PID controller for the measured variables pH, ORP, chlorine, chlorine dioxide, chlorite, ozone, bromine, peracetic acid, hydrogen peroxide, fluoride, dissolved oxygen and conductivity via mA. The sensors for pH and ORP can be directly connected via coaxial cable or using the 4-20 mA sensor input. The controller can bi-directionally control the measured variables, monitor limit values and transmit the measured value via an mA output, e.g. to a PLC Programmable Logic Controller.

The mA output can optionally also be configured as an interference variable output. The controller has two pulse frequency outputs to control two metering pumps (raise and lower). Two output relays can optionally be used as limit value relays or to control motor-driven pumps or solenoid valves. An alarm relay signals the occurrence of a fault. A digital input is used to switch off the control or to process a sample water limit contact by remote control. The impact of temperature on the measurements can be provided by temperature measurement or by manual input. Menu-driven operation is possible in 20 languages.

Your benefits

- Flexibility through free selection of variables from all measured variables.
- Safety through sensor monitoring of pH for glass breakage and line breakage.
- Flexibly upgradable, thanks to subsequent activation option of functions by means of an activation code.
- Various installation options: wall-mounted or installation in a control cabinet.

Field of application

- Measurement and control of water parameters in industrial and process water treatment plants.
- Waste water neutralisation.
- Measurement of the pH value and the disinfection parameters in potable water treatment and in the food and beverage industry.
- Measurement and control of the hygiene parameters in swimming pools.



Transmitter DULCOMETER® DMTa

The compact 2-wire transmitter – the link to the PLC and DULCOMETER®

Technical Data

Measuring range	Type of connection mV:
	pH 0.00 ... 14.00
	ORP - 1,000 ... +1,000 mV
	Type of connection mA:
	Chlorine: 0.00...0.500/2.00/5.00/10.0/20.0/50.0/100.0 ppm
	Chlorine dioxide: 0.00...0.500/2.00/10.0/20.0 ppm
	Chlorite: 0.02...0.50/0.1...2 ppm
	Bromine: 0.02...2.0/0.1...10.0 ppm
	Ozone: 0.00...2.00 ppm
	Hydrogen peroxide, PER1 sensor : 2.0...200.0/20...2,000 ppm
	Peracetic acid: 1...20/10...200/100...2,000 mg/l
	Dissolved oxygen: 0.1...10/0.1...20 ppm
	pH: 0.00...14.00
ORP: 0...+1,000 mV	
Conductivity: 0...20/200/1,000 mS/cm, via mA converter	
Temperature: 0...100 °C via mA converter	
Resolution	pH: 0.00...14.00 ORP: 0...+1,000 mV Amperometric (e. g. chlorine): 0.001/0.01 ppm, 0.01 vol.%
Accuracy	0.5% of the upper measuring range value
Measurement input	SN6 (input resistance > 0.5 x 10 ¹² Ω)
Correction variable	Temperature via Pt 100/Pt 1000
Correction range temp.	0 ... 100 °C
Control characteristic	P/PID control
Control	2-way control
Signal current output	1 x 0/4-20 mA galvanically isolated max. load 450 Ω Adjustable range and allocation (measured variable, correction variable, controlled variable)
Control outputs	2 pulse frequency outputs for metering pump actuation 2 relays (limit value or pulse length)
Alarm relay	250 V ~ 3 A, 700 VA changeover contact
Electrical connection	100 – 230 V, 50/60 Hz, 15 VA
Ambient temperature	-5 ... 50 °C
Enclosure rating	Wall mounting: IP 65 Control panel version: IP 54
Dimensions	Wall mounting: 198 x 200 x 76 mm (WxHxD) Control panel version: 96 x 96 x 145 mm (WxHxD) (D1Cc)
Weight	0.8 kg

■ SYDNEY OFFICE

Unit 4, 4 Narabang Way,
BELROSE NSW 2085
P 02 9450 0995
F 02 9450 0996
sales@prominentfluid.com.au

■ QUEENSLAND OFFICE

Unit 1, 68 Murdoch Circuit,
ACACIA RIDGE QLD 4110
P 07 3213 1900
F 02 9450 0996
pfcqld@prominentfluid.com.au

■ VICTORIA OFFICE

Unit 1/21-22 National Drive
HALLAM VIC 3803
P 03 8795 7430
F 02 9450 0996
pfcvic@prominentfluid.com.au

■ WESTERN AUSTRALIA OFFICE

Office 11, 34 Welshpool Road
WELSHPOOL WA 6106
pfcwa@prominentfluid.com.au